

Specification Sheet

H.S.K. No	DESCRIPTION	Q'TY / UNIT
9018.49-2000	Dental unit chair system "TAURUS Z" (SHINHUNG CO., LTD.)	1 Set

A. Features:

1. For education, diagnosis and patient treatment.
2. Consist of Dental Unit Assembly including cuspidor with suction devices, dental chair, dental operating light, and one stool for dentist, other necessary parts and components.
3. Ergonomic design and total infection control system.
4. It is designed to meet today's stringent infection control demands with ease.
5. A variety of intelligent features assure top-level infection control all instruments including high speed handpieces, 3-way and suction devices and Air-motor are fully autoclavable.
6. Urethane upholstery is seamless and can be cleaned with alcohol.
7. Single-sided backrest hinge allows easy cleaning between seat sections.
8. Flat touch panel controls for easier cleaning.
9. It is controlled by central integrated circuit (I.C)
10. The patient can be relaxed & comfortable in the synchronized movement of chair.

B. Specifications:

1. Dental unit

1) Type: chair mounted unit.

2) Doctor side element includes:

* Instrument holder can accommodate up to 6 instruments

(1) Silicone Tubing for High speed Handpieces -----2 ea.

- ① Available with 2(3)-hole Borden, 4-hole Midwest type connector
Fiber optic connector (optional)
- ② can be disinfected with alcohol

(2) Silicone Tubing for Low speed Handpiece -----1 ea.

- ① Type for Air Motor or Electric Micromotor
- ② can be disinfected with alcohol

(3) Three-way syringe -----1 ea.

- ① DCI 3-way syringe
- ② For air, water and spray
- ③ Tip is removable and easily autoclavable | tip rotatable
- ④ Silicone tube can be disinfected with alcohol.

(4) Bracket table: Ergonomically design -----1 set

- ① Bracket table features a wide range of movement to accommodate dentists' various preference on working position.

- ② Aseptic touch control panel, all grouped in one place for easy access. (some functions optional)

Control panel is controlling:

- Selection of chair Auto position:
Treatment position (1,2), zero position
- Mouth rinse position, Soft memory position
- Manual adjustment of chair position: Seat up/down, backrest up/down
- Memory of chair Auto position
- On/off of Water spray of handpiece (even after instruments go back to holder, last selected mode is memorized)
- On/off of fiber optic of high speed handpiece (even after instruments go back to holder, last selected mode is memorized)
- On /off of Operating light
- Rotation direction of micromotor (clockwise / counterclockwise)
- Manual filling of cup

- ③ Panel can be disinfected with alcohol.

- ④ Standard size (intraoral) x-ray film viewer attached

3) Assistant's side element includes:

(1) Three-way syringe-----1 ea.

- ① DCI 3-way syringe
- ② For air, water and spray
- ③ Tip is removable and easily autoclavable | tip rotatable
- ④ Silicone tube can be disinfected with alcohol.

(2) High volume evacuator -----1 ea.

- ① Lever for controlling suction volume
- ② Aseptic tube can be disinfected with alcohol
- ③ Available for central / air venturi suction
- ④ Easy to disassemble for cleaning. Vacuum valve 135°C autoclavable

(3) Saliva ejector ----- 1 ea.

- ① Lever for controlling suction volume
- ② Aseptic tube can be disinfected with alcohol
- ③ Available for central / air / water venturi suction
- ④ Easy to disassemble for cleaning. Ejector valve 135°C autoclavable

(4) Detachable glass spittoon ----- 1 ea.

- ① Reinforced glass material
- ② Spittoon bowl can be detached for efficient and easy cleaning.
- ③ Spittoon bowl is flushed out with water automatically after cup filling (amount of water for flushing can be adjusted)
- ④ Water warmer is mounted

(5) Assistant's instrument holder

- ① Mounted on extendable arm and swiveling for the best position
- ② Control panel is controlling
 - Selection of chair Auto position:
Treatment position (1,2), zero position.

- Mouth rinse position, Soft memory position
- Manual filling of cup
- Manual flushing of spittoon bowl

(6) Safety stop system:

When Assistant's instrument holder comes into a contact with obstacles during downward movement it will automatically stop and goes up a little to ensure the space for safety.

- (7) Support table (*optional) can be mounted next to control panel to put small instruments during a treatment – in this case, assistant instrument holder cannot be swivel.

2. Dental chair

(1) Type : Oil hydraulic motor drive system

(2) Elevation range : Minimum. : 440 mm \pm 10 mm
Maximum: 740 mm \pm 10 mm

(3) Chair length : max.1820mm

(4) Backrest angle adjustment :

- 0 \pm 2 degree reclines below horizontal for trendelenburg position.
- 75 \pm 2 degree upright for treatment and seated position. :

(5) Tilting :

All movements activated by means of one footswitch located on rear right or left side of chair.

(6) Chair Auto position includes:

Chair provides three programmable automatic positions:

- ① Treatment position 1
- ② Treatment position 2
- ③ Reset (zero) position.

(7) Simple, aseptic chair housing.

(8) Chair has one footswitch for latest in asepsis design

(9) Contour design of chair provides excellent patient support.

(10) Upholstery :

- Seamless urethane material
- Can be disinfected with alcohol.

(11) Headrest (natural movement headrest) :

- The headrest's movement is precisely matched to the movement of the patient neck so there is no uncomfortable sliding of the head.
- Seamless headrest upholstery

- 2 steps articulating headrest

(12) Included electric main switch

(13) Emergency stop function

In case of emergency during chair's movement to Auto position, any of following ways stops chair's movement (totally 6 ways).

- Lifting up assistant holder upward in an instant.
- Turn OFF main switch
- Press any lever of foot controller
- Press main pedal of foot controller
- Press any switch related to chair position on dentist control panel
- Press any switch related to chair position on assistant control panel

3. Other parts

(1) Automatic cup filler -----1 ea.

- Optical type Auto cup filler.
Optical sensor starting cup-filling when cup is placed. Cup filling is made for preset time (filling time can be adjusted)
- Automatic cup filler works with paper or metal cups.

(2) Operating lights -----1 ea.

- 1) The light is post-mounted. Long lasting quartz halogen bulb.
- 2) Auto-light function
 - During chair's movement for Auto position, light OFF automatically
 - On the moment chair gets to Auto treatment position, light ON automatically
- 3) Detachable, autoclavable light handle.
- 4) Light intensity: Max. 24,000 \pm 2,000 Lux
Low: 13,000 Lux High: 24,000 Lux
- 5) Color temperature: 4,200 \pm 2,000 Kelvin
- 6) Knob for adjusting Light pattern is mounted on rear side of light head.

(3) Foot controllers ----- 1 ea.

- All manual movements of chair are also controllable with foot controller.
(Chair elevation range, Backrest angle)
- Control on-off for ultrasonic scaler
- Control handpiece speed.

(4) Operating stool (Optional)

- 1) 'Sante stool' (for dentist)
 - Gas cylinder type, ergonomic type seat
 - Backrest can be tilted ergonomically (tilting on/off and tension selectable by lever)
 - Seamless upholstery: Easy cleaning with alcohol.
 - Smooth operation 5 castors.
 - Elevation Range : Min.460mm - Max.570mm
 - Width: 540mm
 - Weight: 14kg

2) 'Taurus stool' (for assistant)

- Gas cylinder type, round type seat
- Seamless upholstery: Easy cleaning with alcohol.
- Smooth operation 5 castors.
- Elevation Range : Min.470mm - Max.620mm
- Width: 570mm
- Weight: 13.7kg

4. Electric Power : AC 220V, 60Hz. (Adjustable to local voltage)

5. Power consumption: 1.5 kVA